

**EC Certificate**  
**Directive 93/42/EEC Annex V**  
**Production Quality Assurance**  
**Medical Devices**

**Registration No.:** DD 60145760 0001

**Report No.:** 16806100 007

**Manufacturer:** XINLE HUABAO MEDICAL  
PRODUCTS CO., LTD.  
Dongguan, Cheng'an Town  
050701 Xinle City, Hebei Province  
P.R. China

**Products:** For the Following Medical Devices the Scope Covers only  
the Aspects of Manufacture Concerned with Securing and  
Maintaining Sterile Conditions

(See Attachment for Products included)

Replaces Approval, Registration No.: DD 60110792 0001

**Expiry Date:** 2024-05-26

The Notified Body hereby declares that the requirements of Annex V of the directive 93/42/EEC have been met for the listed products. The above named manufacturer has established and applies a quality assurance system, which is subject to periodic surveillance, defined by Annex V, section 4 of the aforementioned directive. For placing on the market of class IIb and class III devices covered by this certificate an EC type-examination certificate according to Annex III is required.

**Effective Date:** 2020-01-19

**Date:** 2020-01-19

Notified Body  
  
Jing Zhang

**TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg**  
TÜV Rheinland LGA Products GmbH is a Notified Body according to Directive 93/42/EEC  
concerning medical devices with the identification number 0197.

**TÜV Rheinland**  
**LGA Products GmbH**  
**Tillystraße 2, 90431 Nürnberg**

Doc. 1/1, Rev. 0

**Attachment to  
Certificate**

**Registration No.:** DD 60145760 0001  
**Report No.:** 16806100 007

**Manufacturer:** XINLE HUABAO MEDICAL  
PRODUCTS CO., LTD.  
Dongguan, Cheng'an Town  
050701 Xinle City, Hebei Province  
P.R. China

**Products:**

- Sterile Non-woven Products (Surgical Gowns, Surgical Drapes, Isolation Gowns, Coveralls, Face Masks)
- Sterile Non-woven and PE Composited Products (Surgical Gowns, Surgical Drapes, Isolation Gowns, Coveralls)

**Date:** 2020-01-19

**Notified Body**



**Business Stream Products**  
**Certification Department**



**TÜVRheinland®**  
**LGA**

**Precisely Right.**

TÜV Rheinland LGA Products GmbH · 90431 Nürnberg

XINLE HUABAO MEDICAL  
PRODUCTS CO., LTD.  
Dongguan, Cheng'an Town  
050701 XINLE CITY, HEBEI PROVINCE  
P.R. CHINA

**Contact**

Tel. +49 911 655-5225  
Mail [service@de.tuv.com](mailto:service@de.tuv.com)

Date January 19, 2020

**Application for : QMS Produktion, Anhang V MDD**  
Certificate No. : DD 60145760 Sheet 0001  
Device : Only for QM-System audit  
Test requirement : Richtlinie 93/42/EWG

Dear Madame or Sir,

Enclosed please find the new certificate No. DD 60145760 0001 replacing the previous certificate.

With effective date of the new certificate, the previous certificate (number see new certificate) becomes invalid.

Kind regards

Certification body

  
Jing Zhang

Test sample: no, documentation available

TÜV Rheinland  
LGA Products GmbH

Tillystraße 2  
90431 Nürnberg

Tel. +49 911 655-5225  
Fax +49 911 655-5226  
Mail [service@de.tuv.com](mailto:service@de.tuv.com)  
Web [www.tuv.com/safety](http://www.tuv.com/safety)

Board of Management

Dipl.-Ing.  
Jörg Mähler, Spokesman

Dipl.-Kfm.  
Dr. Jörg Schlösser

Chairman of the  
Supervisory Board

Dipl.-Ing.  
Ralf Scheller

Nuremberg HRB 26013  
VAT No.: DE 811835490

# DICHIARAZIONE DI CONFORMITÀ

(ALLEGATO VII DIRETTIVA 93/42/CEE)



Identificativo di registrazione Repertorio BD/RDM: 2027953

*Fabbricante:*

**Xinle Huabao Medical Products Co., Ltd.**

**Dongguan, Cheng, an Town, Xinle City, Hebei Province, 050701, China**

*Mandatario Europeo:*

**Wellkang Ltd. Suite B, 29 Harley Street, London W1G 9QR, UK**

***Del prodotto:***

*Denominazione*

**CAMICE MEDICO monouso**

*Tipologia / Classe dispositivo*

**Classe I - Non sterile**

*Marca*

**XINLE HUABAO MEDICAL PRODUCTS**

*Categoria*

**B [B004] S - M - L - XL - XXL**

*Anno di Costruzione*

**2020**

*Lotto*

**Il prodotto sopraindicato è conforme alle disposizioni delle seguenti direttive e successivi emendamenti:**

**93/42/CEE Direttiva Dispositivi Medici e modificata dalla Direttiva 2007/47/EC.**

**Dichiara inoltre che il prodotto è conforme ai seguenti standard:**

**EN 13795:2019 Camici monouso Classe I ad uso medico – Requisiti e metodi di prova.**

Procedura di valutazione

della conformità: **Modulo A (CE Dichiarazione di conformità (allegato VII) + fascicoli tecnici)**

Roma, 28/10/2020

*Firma*

*Fabbricante:* 刘敏奇

新乐华宝医疗用品有限公司  
XINLEHUABAO MEDICAL PRODUCTS CO., LTD  
刘敏奇  
LIU MIN QI

## Test Report

SL52025269537001TX

Date: July 07, 2020

Page 1 of 11

XINLE HUABAO PLASTIC PRODUCTS CO., LTD.  
NO.210, NANHUAN ROAD, XINLE CITY, HEBEI PROVINCE

The following sample(s) was/were submitted and identified on behalf of the client as:

Sample Description : (A) The Disposable Isolation Gown is made from nonwoven and PE composited material, and consists of a front and back cut piece (including two sleeves, two collar bindings and two belts). The product is suitable for general isolation in out-patient, sickroom and inspection room of medical institution.

SGS Internal Ref No. : SL120012511014TX

Roll/ Lot No. : 20200520DDR

Sample Receiving Date : Jun 04, 2020

Testing Period : Jun 04, 2020 – Jul 07, 2020

Test Result(s) : Unless otherwise stated the results shown in this test report refer only to the sample(s) tested, for further details, please refer to the following page(s).

Test Performed : Selected test(s) as requested by applicant



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)

3<sup>rd</sup> Building, No. 889, Yishan Road, Xuhui District Shanghai, China 200233  
中国·上海·徐汇区宜山路889号3号楼 邮编: 200233

t (86-21) 61402666 f (86-21) 64958763  
t (86-21) 61402666 f (86-21) 64958763

[www.sgsgroup.com.cn](http://www.sgsgroup.com.cn)  
e [sgs.china@sgs.com](mailto:sgs.china@sgs.com)




**Comment:**

<b>Surgical Clothing and Drapes - Requirements and Test Methods - Part 1: Surgical Drapes and Gowns</b> (EN 13795-1:2019)	(A)
Clause 4 Bursting strength-Dry	M
Clause 4 Bursting strength-Wet	M
Clause 4 Tensile strength-Dry	M
Clause 4 Tensile strength-Wet	M
Clause 4 Liquid penetration-Water Resistance	M
Clause 4 Cleanliness microbial/Bioburden	M
Clause 4 Microbial Penetration-Dry	M
Clause 4 Microbial penetration-Wet	M
Clause 4 Particle release	M

Remark: M=Meet EN 13795-1:2019 standard performance requirement

Signed for and on behalf of  
SGS-CSTC Standards Technical Services (Shanghai) Co., Ltd Testing Center



Sara Guo (Account Executive)



Dongjing Liu (Authorized Signatory)



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing /inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)

3<sup>rd</sup> Building, No.889, Yishan Road, Xuhui District Shanghai, China 200233  
中国·上海·徐汇区宜山路889号3号楼 邮编: 200233

t (86-21) 61402666 f (86-21) 64958763  
t (86-21) 61402666 f (86-21) 64958763

[www.sgsgroup.com.cn](http://www.sgsgroup.com.cn)  
e [sgs.china@sgs.com](mailto:sgs.china@sgs.com)

Test Result

**Surgical Clothing and Drapes - Requirements and Test Methods - Part 1: Surgical Drapes and Gowns**  
(EN 13795-1:2019)

**Clause 4 Bursting strength-Dry\***

(EN ISO 13938-1: 1999; Dry state; Test area: 10cm<sup>2</sup>)

Material

Sample	Measured value (kPa)	Requirement (kPa)	Conclusion
	Face		
1	117	≥40  (Surgical gown: standard performance critical product area)  EN 13795-1:2019	Pass
2	97.0		
3	104		
4	112		
5	108		

Sleeve seam

Sample	Measured value (kPa)	Requirement (kPa)	Conclusion
	Face		
1	134	≥40  (Surgical gown: standard performance critical product area)  EN 13795-1:2019	Pass
2	128		
3	132		
4	131		
5	130		



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)

SGS-CTC Standards Technical Services (Shanghai) Co., Ltd.  
Testing Center for Textiles

3<sup>rd</sup> Building, No. 889, Yishan Road, Xuhui District Shanghai, China 200233  
中国·上海·徐汇区宜山路889号3号楼 邮编: 200233

t (86-21) 61402666 f (86-21) 64958763  
t (86-21) 61402666 f (86-21) 64958763

[www.sgsgroup.com.cn](http://www.sgsgroup.com.cn)  
e [sgs.china@sgs.com](mailto:sgs.china@sgs.com)

**Clause 4 Bursting strength-Wet\***

(EN ISO 13938-1: 1999; Dry state; Test area: 10cm<sup>2</sup>)

**Material**

Sample	Measured value (kPa)	Requirement (kPa)	Conclusion
	Face		
1	69.4	≥40  (Surgical gown: standard performance critical product area)  EN 13795-1:2019	Pass
2	99.4		
3	87.3		
4	74.9		
5	92.3		

**Sleeve seam**

Sample	Measured value (kPa)	Requirement (kPa)	Conclusion
	Face		
1	134	≥40  (Surgical gown: standard performance critical product area)  EN 13795-1:2019	Pass
2	126		
3	128		
4	133		
5	131		



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)

3<sup>rd</sup> Building, No.889, Yishan Road, Xuhui District Shanghai, China 200233  
中国·上海·徐汇区宜山路889号3号楼 邮编: 200233

t (86-21) 61402666 f (86-21) 64958763  
t (86-21) 61402666 f (86-21) 64958763

[www.sgsgroup.com.cn](http://www.sgsgroup.com.cn)  
e [sgs.china@sgs.com](mailto:sgs.china@sgs.com)



## Clause 4 Tensile strength-Dry

(EN 29073-3:1992; Dry state; The distance between the clamps: 200mm; Rate: 100mm/min)

(EN 29073-3:1992 eqv ISO 9073-3:1989)

### Material

Sample	Length (N)	Width (N)	Requirement (N)	Conclusion
1	48.1	70.9	$\geq 20$  (Surgical gown: standard performance critical product area)  EN 13795-1:2019	Pass
2	54.0	78.6		
3	55.7	72.5		
4	52.8	68.6		
5	49.7	75.5		

### Sleeve seam

Sample	(N)	Requirement (N)	Conclusion
1	38.0	$\geq 20$  (Surgical gown: standard performance critical product area)  EN 13795-1:2019	Pass
2	34.2		
3	34.0		
4	36.8		
5	38.3		



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)

3<sup>rd</sup> Building, No. 889, Yishan Road, Xuhui District Shanghai, China 200233  
中国·上海·徐汇区宜山路889号3号楼 邮编: 200233

t (86-21) 61402666 f (86-21) 64958763  
t (86-21) 61402666 f (86-21) 64958763

[www.sgs.com](http://www.sgs.com)  
e [sgs.china@sgs.com](mailto:sgs.china@sgs.com)

## Clause 4 Tensile strength-Wet

(EN 29073-3: 1992; Wet state. The distance between the clamps: 200mm; Rate: 100mm/min)

(EN 29073-3:1992 eqv ISO 9073-3:1989)

### Material

Sample	Length (N)	Width (N)	Requirement (N)	Conclusion
1	47.3	75.7	≥20  (Surgical gown: standard performance critical product area)  EN 13795-1:2019	Pass
2	48.8	69.0		
3	46.8	79.1		
4	45.1	73.8		
5	47.2	73.2		

### Sleeve seam

Sample	(N)	Requirement (N)	Conclusion
1	41.2	≥20  (Surgical gown: standard performance critical product area)  EN 13795-1:2019	Pass
2	38.4		
3	37.4		
4	37.9		
5	37.4		



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)

SGS-CTC Standards Technical Services (Shanghai) Co., Ltd.  
Testing Center for Textiles

3<sup>rd</sup> Building, No. 889, Yishan Road, Xuhui District Shanghai, China 200233  
中国·上海·徐汇区宜山路889号3号楼 邮编: 200233

t (86-21) 61402666 f (86-21) 64958763  
t (86-21) 61402666 f (86-21) 64958763

[www.sgsgroup.com.cn](http://www.sgsgroup.com.cn)  
e [sgs.china@sgs.com](mailto:sgs.china@sgs.com)

## Clause 4 Liquid penetration-Water Resistance

(EN ISO 811: 2018; Rate of increase of water pressure: 10cm/min; temp. of distilled water: 20.0°C; Face side to water)

### Material

Sample	Measured value (cmH <sub>2</sub> O)	Requirement (cmH <sub>2</sub> O)	Conclusion
1	> 20	≥20 (Surgical gown: standard performance critical product area) EN 13795-1:2019	Pass
2	> 20		
3	> 20		
4	> 20		
5	> 20		

### Sleeve seam

Sample	Measured value (cmH <sub>2</sub> O)	Requirement (cmH <sub>2</sub> O)	Conclusion
1	> 20	≥20 (Surgical gown: standard performance critical product area) EN 13795-1:2019	Pass
2	> 20		
3	> 20		
4	> 20		
5	> 20		

## Clause 4 Cleanliness microbial/Bioburden

(EN ISO 11737-1: 2018; Membrane Filtration Method)

Sample	total plate count (CFU/100cm <sup>2</sup> )	Requirement (CFU/100cm <sup>2</sup> )	Conclusion
1	30	≤300 (Surgical gown: standard performance critical product area) EN 13795-1:2019	Pass
2	240		
3	66		
4	33		
5	270		



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)

3<sup>rd</sup> Building, No. 889, Yishan Road, Xuhui District Shanghai, China 200233  
中国·上海·徐汇区宜山路889号3号楼 邮编: 200233

t (86-21) 61402666 f (86-21) 64958763  
t (86-21) 61402666 f (86-21) 64958763

[www.sgsgroup.com.cn](http://www.sgsgroup.com.cn)  
e [sgs.china@sgs.com](mailto:sgs.china@sgs.com)

## Clause 4 Microbial Penetration-Dry\*

(EN ISO 22612:2005, The fourth generation of spores of bacillus subtilis ATCC 9372, The concentration of the spores:  $1.8 \times 10^8$  CFU/g talcum powder, Sample: 12, Vibration frequency: 20800 times/min, Vibration time: 30min)

### Material

Sample	Measured value (CFU)	Requirement (CFU)	Conclusion
1	0	$\leq 300$  (Surgical gown: standard performance less critical product area)  EN 13795-1:2019	Pass
2	0		
3	0		
4	0		
5	0		
6	0		
7	0		
8	0		
9	0		
10	0		

### Sleeve seam

Sample	Measured value (CFU)	Requirement (CFU)	Conclusion
1	0	$\leq 300$  (Surgical gown: standard performance less critical product area)  EN 13795-1:2019	Pass
2	0		
3	0		
4	0		
5	0		
6	0		
7	0		
8	0		
9	0		
10	0		



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)

SGS-CSI Standards Technical Services (Shanghai) Co., Ltd.  
Testing Center for Textiles

3<sup>rd</sup> Building, No. 889, Yishan Road, Xuhui District Shanghai, China 200233  
中国·上海·徐汇区宜山路889号3号楼 邮编: 200233

t (86-21) 61402666 f (86-21) 64958763  
t (86-21) 61402666 f (86-21) 64958763

[www.sgsgroup.com.cn](http://www.sgsgroup.com.cn)  
e [sgs.china@sgs.com](mailto:sgs.china@sgs.com)

## Clause 4 Microbial penetration-Wet\*

(EN ISO 22610:2006, Temp: 24.5°C, RH: 56.0%, The distance of agar to plates brim: 3mm, Carrier material: 30µm polyurethane (PU))

### Material

Sample	Barrier index	Requirement Barrier index	Conclusion
1	6.0	$\geq 2.8$ (Surgical gown: standard performance critical product area) EN 13795-1:2019	Pass
2	6.0		
3	6.0		
4	6.0		
5	6.0		

### Sleeve seam

Sample	Barrier index	Requirement Barrier index	Conclusion
1	6.0	$\geq 2.8$ (Surgical gown: standard performance critical product area) EN 13795-1:2019	Pass
2	6.0		
3	6.0		
4	6.0		
5	6.0		



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)

SGS-CSI Standards Technical Services (Shanghai) Co., Ltd.  
Testing Center for Textiles

3<sup>rd</sup> Building, No. 889, Yishan Road, Xuhui District Shanghai, China 200233  
中国·上海·徐汇区宜山路889号3号楼 邮编: 200233

t (86-21) 61402666 f (86-21) 64958763  
t (86-21) 61402666 f (86-21) 64958763

[www.sgsgroup.com.cn](http://www.sgsgroup.com.cn)  
e [sgs.china@sgs.com](mailto:sgs.china@sgs.com)



**Clause 4 Particle release\***

(EN ISO 9073-10: 2004; Size of particles counted: 3µm-25µm)

(EN ISO 9073-10:2004 idt ISO 9073-10:2003)

**Material**

Size of particles counted (μm)	Sample		Measured value Coefficient of linting log <sub>10</sub>	Requirement Coefficient of linting log <sub>10</sub>	Conclusion
3~25	A: Face	1	2.6	≤4.0  (Surgical gown: standard performance critical product area)  EN 13795-1:2019	Pass
		2	2.3		
		3	2.5		
		4	2.6		
		5	2.3		
	A: Back	1	2.6		
		2	2.6		
		3	2.6		
		4	2.7		
		5	2.6		

**Sleeve seam**

Size of particles counted (μm)	Sample		Measured value Coefficient of linting log <sub>10</sub>	Requirement Coefficient of linting log <sub>10</sub>	Conclusion
3~25	A: Face	1	2.3	≤4.0  (Surgical gown: standard performance critical product area)  EN 13795-1:2019	Pass
		2	2.0		
		3	2.2		
		4	2.3		
		5	2.0		
	A: Back	1	2.9		
		2	2.8		
		3	2.9		
		4	3.0		
		5	2.8		

\*This test standard is not within the accredited scope in SGS Shanghai testing centre, it is carried out by external laboratory accredited by CNAS (China National Accreditation Service for Conformity Assessment).



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Document.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)

3<sup>rd</sup> Building, No.889, Yishan Road, Xuhui District Shanghai, China 200233  
中国·上海·徐汇区宜山路889号3号楼 邮编: 200233

t (86-21) 61402666  
t (86-21) 61402666

f (86-21) 64958763  
f (86-21) 64958763

[www.sgsgroup.com.cn](http://www.sgsgroup.com.cn)  
e [sgs.china@sgs.com](mailto:sgs.china@sgs.com)

Sample Photo



Face

Back

The statement of conformity in this test report is only based on measured values by the laboratory and does not take their uncertainties into consideration.

\*\*\*End of Report\*\*\*



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone: (86-755) 8307 1443, or email: [CN.Doccheck@sgs.com](mailto:CN.Doccheck@sgs.com)

3<sup>rd</sup> Building, No. 889, Yishan Road, Xuhui District Shanghai, China 200233  
中国·上海·徐汇区宜山路889号3号楼 邮编: 200233

t (86-21) 61402666 f (86-21) 64958763  
t (86-21) 61402666 f (86-21) 64958763

[www.sgs.com](http://www.sgs.com)  
e [sgs.china@sgs.com](mailto:sgs.china@sgs.com)

**Business Stream Products**  
**Certification Department**



**TÜVRheinland®**  
**LGA**

**Precisely Right.**

TÜV Rheinland LGA Products GmbH · 90431 Nürnberg

XINLE HUABAO MEDICAL  
PRODUCTS CO., LTD.  
Dongguan, Cheng'an Town  
050701 XINLE CITY, HEBEI PROVINCE  
CHINA

**Contact**

Tel. +49 911 655-5225  
Mail [service@de.tuv.com](mailto:service@de.tuv.com)

Date October 31, 2019

**Application for : QMS**  
Certificate No. : SX 60143175 Sheet 0001  
Device : Only for QM-System audit  
Test requirement : EN ISO 13485:2016

Dear Madame or Sir,

Enclosed please find the  
new certificate No. SX 60143175 0001  
replacing the previous certificate.

Kind regards

Certification body

Jing Zhang

Test sample: no, documentation available

TÜV Rheinland  
LGA Products GmbH

Tillystraße 2  
90431 Nürnberg

Tel. +49 911 655-5225  
Fax +49 911 655-5226  
Mail [service@de.tuv.com](mailto:service@de.tuv.com)  
Web [www.tuv.com/safety](http://www.tuv.com/safety)

Board of Management

Dipl.-Ing.  
Jörg Mähler, Spokesman

Dipl.-Kfm.  
Dr. Jörg Schlösser

Chairman of the  
Supervisory Board

Dipl.-Ing.  
Ralf Scheller

Nuremberg HRB 26013  
VAT No.: DE 811835490

**TÜV Rheinland**  
**LGA Products GmbH**  
**Tillystraße 2, 90431 Nürnberg**

**Attachment to  
Certificate**

**Registration No.:** SX 60143175 0001  
**Report No.:** 16806100 005

**Organization:** XINLE HUABAO MEDICAL  
PRODUCTS CO., LTD.  
Dongguan, Cheng'an Town  
050701 Xinle City, Hebei Province  
China

**Scope:**


**Products:**

- Non-woven Products
- Non-woven and PE composited Products
- PE Products
- Paper Products
- PE and Paper composited Products

**Certification Body**



**Date: 2019-10-31**



# Certificate

The Certification Body of  
TÜV Rheinland LGA Products GmbH

hereby certifies that the organization

**XINLE HUABAO MEDICAL  
PRODUCTS CO., LTD.**  
Dongguan, Cheng'an Town  
050701 Xinle City, Hebei Province  
China

has established and applies a quality management system for medical devices  
for the following scope:

**Manufacture and Distribution of Disposable Medical Devices**  
**(see attachment for products included)**

Proof has been furnished that the requirements specified in

**EN ISO 13485:2016**

are fulfilled. The quality management system is subject to yearly surveillance.

Effective Date: 2019-10-31  
Certificate Registration No.: SX 60143175 0001  
An audit was performed. Report No.: 16806100 005  
This Certificate is valid until: 2022-07-07

Certification Body



Date 2019-10-31



**TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg**  
Tel.: +49 221 806-1371 Fax: +49 221 806-3935 e-mail: cert-validity@de.tuv.com <http://www.tuv.com/safety>